

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

This document provides general technical guidance for persons importing Medical Devices on what declaration information to report via the Automated Commercial Environment (ACE). It is not intended to cover all specific scenarios. Please refer to the FDA Supplemental Guide to the CBP and Trade Automated Interface Requirements (CATAIR) for information on how to make electronic filings. The document is not an official statement of the Federal government and does not create any requirements; rather it explains what information to input in accordance with the CBP pilots for FDA Medical Devices.

Medical Devices

- Medical devices range from simple tongue depressors and bedpans to complex programmable pacemakers with micro-chip technology and laser surgical devices.
- Medical devices include in vitro diagnostic products, such as general purpose lab equipment, reagents, and test kits.
- Certain electronic [radiation emitting products](#) with medical application and claims meet the definition of medical device. Examples include diagnostic ultrasound products, x-ray machines and medical lasers. FDA's Center for Medical Devices and Radiological Health regulates a range of products from microwaves to DVD drives.
- If a product is labeled, promoted or used in a manner that meets the following definition in section 201(h) of the Federal Food Drug & Cosmetic (FD&C) Act it will be regulated by the [Food and Drug Administration \(FDA\)](#) as a medical device and is subject to premarketing and postmarketing regulatory controls.

Medical Device Definition

- "an instrument, apparatus, implement, machine, contrivance, implant, in vitro reagent, or other similar or related article, including a component part, or accessory which is:
 - Recognized in the official National Formulary, or the United States Pharmacopoeia, or any supplement to them
 - Intended for use in the diagnosis of disease or other conditions, or in the cure, mitigation, treatment, or prevention of disease, in man or other animals, or
 - Intended to affect the structure or any function of the body of man or other animals, and which does not achieve its primary intended purposes through chemical action within or on the body of man or other animals

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

and which is not dependent upon being metabolized for the achievement of any of its primary intended purposes."

- Some non-medical purposes are also regulated as medical devices due to safety concerns.
- The intended use of a device or certain claims made on the labeling could also make it a medical device.

When transmitting an FDA Medical Device, the PGA records that are utilized are:

Description
Commercial Description
Government Agency Code (FDA)
Government Program Code (DEV)
Government Agency Processing Code (RED, NED)
Intended Use Code/Intended Use Description
Product Code
Country of Production/Manufacturing
Trade/Brand Name
Invoice / Item Description
Manufacturer Name, Address, FEI
Importer of Record Name, Address, FEI
Shipper Name, Address, FEI
Delivered to Party Name, Address, FEI
Device Initial Importer Name, Address, FEI
Point of Contact Name, Email Address
Affirmations of Compliance
General Remarks
Line Value
All Levels of Packaging (PCS must be base unit)
Anticipated Arrival Date and Location
Disclaimer (If disclaiming, FDA does not require the above information)

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Commercial Description: The commercial description of the product. For example, PEDIATRIC TOURNIQUET CUFF SET.

Government Agency Code: FDA

Government Agency Program code for FDA Medical Device PGA Message Sets:

Government Agency Program Code	Description
DEV	Medical Devices

Government Agency Processing Code:

Government Agency Processing Code	Description
RED	Radiation Emitting Devices
NED	Non-Radiation Emitting Devices

Intended Use Code

For Medical Devices, only one of the following Intended Use Codes may be entered:

Intended Use Description

This field is used to describe the Intended Use such as 'Sample devices', 'Return shipment', etc

Intended Use Code	Intended Use Definition	Relevant Medical Device Import Scenarios
081.001	For Human Medical Use as a Medical Device	<ul style="list-style-type: none"> • Standard import of a medical device, accessories, or components regulated as a finished device • Import of refurbished device • Import of a reprocessed device
081.002	For Human Medical Use as a Medical Device for Domestic Refurbishing	
081.003	For Human Medical Use as Medical Device—domestically manufactured device that is part of a medical device convenience kit	
081.004	For Human Medical Use as a Medical Device – foreign manufactured device that is part of a medical device convenience kit	
081.005	Importation of a device constituent part (finished device) for use in a medical product regulated under a drug (CDER) application type (e.g., for use in an NDA/ANDA/BLA drug- device combination product).	
081.006	Import of a medical device under enforcement discretion provisions per final guidance: http://www.fda.gov/ucm/groups/fdagov-public/@fdagov-meddev-gen/documents/document/ucm401996.pdf	Applicable product codes for CBP business rule: 800 -- UG; 86N -- FF 86N -- FG; 80N -- XQ 90L -- MB; 90L -- MD
100.000	For Personal Use as a Non-Food Product – for personal use as a medical device	
110.000	For Public Exhibition or Display as a Non-Food Product	<ul style="list-style-type: none"> • Includes import of device for trade show

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Intended Use Code	Intended Use Definition	Relevant Medical Device Import Scenarios
140.000	For Charitable Organization Use as a Non-Food Product	
155.010	Component for further manufacturing into a finished medical device	
155.011	Importation of a device component for use in a medical product regulated under a drug (CDER) application type (e.g., for use in an NDA/ANDA/BLA drug-device combination product).	
170.000	For Repair of a Non-Food Product	<ul style="list-style-type: none"> • Repair of medical device and re-exportation
180.010	For Research and Development as a Non-Food Product - For research and development as a medical device	<ul style="list-style-type: none"> • Import of research or investigational use in vitro diagnostic device
180.014	For Research and Development as a Non-Food Product – for bench testing or nonclinical research use	<ul style="list-style-type: none"> • Import of a device for non-clinical use/bench testing • Import of device sample for customer evaluation
180.015	For Research and Development as a Non-Food Product – import of a medical device for clinical investigational use	
920.001	Import of a device that is US goods returned for refund/overstock (to manufacturer)	
920.002	Import of a device that is US goods returned for sale to a third party	
940.000	Import of a Compassionate Use/Emergency Use Device	
950.001	Import of a single-use device for domestic reprocessing	
950.002	Import of a multi-use device for domestic reprocessing	
970.000	Import for Export	<ul style="list-style-type: none"> • Import of a medical device for further processing and re-exportation • Import of medical device or accessory for further manufacturing into an export only medical device
970.002	Import for Export	<ul style="list-style-type: none"> • Import of a medical device component for further manufacturing into an export only medical device

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Product Code

Only one Product Code Number per product is allowed. For components, use the product code of the most likely device the component will be incorporated into if not known.

FDA Product Code Builder Tutorial:

<http://www.accessdata.fda.gov/scripts/ora/pcb/tutorial/tutorial.cfm>

Product Code Must be equal to 7 characters

FDA Product Code Structure:

Position	1-2	3	4	5	6-7
Name	Industry Code (N)	Class Code (A)	Sub Class Code (A or '-')	Process Identification Code - PIC (A or '-')	Product (AN)

Country of Production/Manufacturing

Country of production or source is required for Medical Devices. (ISO Country Code)

<http://www.cbp.gov/document/guidance/appendix-c-iso-country-codes>

Trade/Brand Name

Trade/Brand Name of the Medical Device. For example, Zimmer Reusable Tourniquet Cuff

Invoice/Item Description

The medical device detail description. NOT product code description.

List of Entity Role codes applicable to FDA Medical Device Message Sets:

Data Element	Code	Description
Entity Role Codes	MF	Manufacturer of goods
	DEQ	Shipper
	FD1	FDA Importer 1 (Importer of Record)
	DII	Device Initial Importer
	DP	Delivered To Party

Entity Information:

Manufacturer of Goods (MF) – Medical Device

Manufacturer means any person who manufactures, prepares, propagates, compounds, assembles, or processes a device by chemical, physical, biological, or other procedure. **The term includes any person who either:**

(1) Repackages or otherwise changes the container, wrapper, or labeling of a device in furtherance of the distribution of the device from the original place of manufacture;

(2) Initiates specifications for devices that are manufactured by a second party for subsequent distribution by the person initiating the specifications;

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

(3) Manufactures components or accessories that are devices that are ready to be used and are intended to be commercially distributed and intended to be used as is, or are processed by a licensed practitioner or other qualified person to meet the needs of a particular patient; or

(4) Is the U.S. agent of a foreign manufacturer.

Shipper (DEQ)

The FDA Shipper is the actual shipper of the product. This can typically be determined from the freight bill or bill of lading, etc. The FDA Shipper may be the same entity as the invoicing party.

FDA Importer (FD1)

Importer means, for purposes of this part, a company or individual in the United States that is an owner, consignee, or recipient, even if not the initial owner, consignee, or recipient, of the foreign establishment's device that is imported into the United States. An importer does not include the consumer or patient who ultimately purchases, receives, or uses the device, unless the foreign establishment ships the device directly to the consumer or patient.

Device Initial Importers (DII)

The initial importer of the device must register its establishment with FDA. An initial importer is any importer who furthers the marketing of a device from a foreign manufacturer to the person who makes the final delivery or sale of the device to the ultimate consumer or user, but does not repackage, or otherwise change the container, wrapper, or labeling of the device or device package. Registration information can be found under [Establishment Registration](#).

Initial importers are also subject to [Medical Device Reporting](#) (MDR) under 21 CFR 803, [Reports of Corrections and Removals](#) under 21 CFR 806, and [Medical Device Tracking](#) under 21 CFR 821, if applicable. Under the MDR regulations importers are required to report incidents in which a device may have caused or contributed to a death or serious injury as well as report certain malfunctions. The importers must maintain an MDR event file for each adverse event. All product complaints (MDR and nonMDR events) must be forwarded to the manufacturer. Under Medical Device Tracking requirements, certain devices must be tracked through the distribution chain.

Delivered To Party (DP)

Site where the goods are being delivered. It represents the facility to physically receive the goods after arrival in the US.

Entity Name and Entity Address (Entity State/Province - Populated ONLY if US or Canada based entities).

FDA requires Entity Name and Entity Address. Additionally, FDA/CDRH prefers to use FEI numbers for identifying the Entity; IF FEI is not available THEN DUNS.

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Registration Number

FDA/CDRH prefers to use FEI numbers for identifying the Entity for Medical Devices;

For devices the vast majority of registration numbers (DEV) are FEIs

IF FEI is not available then provide a DUNS.

(FEI) THEN Entity Number MUST BE Length from 4 to 10

(DUNS) THEN Entity Number MUST BE Length = 9

Point of Contact

Data Element	Code	Description
Entity Role Codes	PK	Point of Contact

Included in this record are the Individual Name, Telephone Number, Fax Number, and Email address. A typical example will be a POC is needed for the Filer. Providing the broker's name and phone number will assist FDA in contacting the appropriate point of contact if there are any discrepancies with the line/entry.

Individual Name
Telephone Number of the Individual
Email Address or Fax Number for the Individual

Affirmations of Compliance

The list of Affirmation of Compliance (AoC) codes for FDA-Medical Devices Message Sets is below followed by the scenarios when the AofC' s should be provided:

The FDA Affirmation of Compliance Codes and their descriptions are listed in the Appendix PGA (Food & Drug Affirmation of Compliance, FDA Affirmation of Compliance Codes) of ACE ABI CATAIR publication.

N=Numeric digits; X=Alphanumeric

Code	Description	Qualifier
PM#	Device Premarket Number <ul style="list-style-type: none"> • Premarket Approval (PMA) number (i.e. Pxxxxxx) • Device Premarket Notification Number (PMN) ((510(k)) (i.e. Kxxxxxx) PMN or PMA number database found at: http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPMN/pmn.cfm http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfpma/pma.cfm <ul style="list-style-type: none"> • Product Development Protocols (PDP) number • Humanitarian Device Exemption (HDE) number 	Any of the following: P+6N; N+4N, 5N, or 6N; D+6N; H+6N; K+6N; DEN+6N
DDM	Device Domestic Manufacturer	1 - 10N
DEV	Device Foreign Manufacturer Registration Number The qualifier for this code should be the device registration number issued by CDRH (Center for Devices and Radiological Health) for the firm manufacturing the product. Note: The DEV should always be associated with the foreign manufacturer.	1 - 10N

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

DFE	Device Foreign Exporter Registration Number	1 - 10N
DI	Device Identifier	6-23X
CPT	Component Identifier	Indicator only
IFE	Import For Export	Indicator only
IDE	Investigational Device Exemption Number	G+6N OR "NSR"
IRC	Device Impact Resistance Lens Certification (Drop Ball Test) <ul style="list-style-type: none"> ▪ This code is used to certify that the filer has on hand the test results or a certificate that shows that the product on the FDA line has met the standards for impact resistance Lens.(sun glasses etc.) ▪ Note: Each shipment must have its own test results. 	Indicator only
KIT	Device Imported Kit of Finished Device	Indicator only
LST	Device Listing Number The qualifier for this code should be the device listing number issued by CDRH for the product identified in the FDA Line	A+6N; B+6N; C+6N; D+6N; E+6N; L+6N; Q+6N; R+6N
LWC	Electrode Lead Wire Or Patient Cable <ul style="list-style-type: none"> • This Code should be used when importing electrode lead wires, patient cables, or devices that use them. • The Affirmation indicates that the device does not contain any pre-wired electrodes, electrode lead wires, or patient (transducer) cables 	Indicator Only

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

The table below shows which Affirmations of Compliance are Mandatory (M), Conditional (C) or Optional (O) based on the Intended Use Code/Import Scenario:

Intended Use (see PG01 for definitions)	Import Scenarios	Mandatory Affirmations	Conditional ¹ Affirmations	Optional Affirmations
081.001	<ul style="list-style-type: none"> Standard import of device, accessories, or components regulated as a finished device Import of refurbished device Import of a reprocessed device 	DEV, DFE, LST	DI, IRC, LWC, PM#	
081.002*	Import of a device for domestic refurbishing	DEV, DFE, LST	DI, IRC, LWC, PM#	
081.003	domestically manufactured device that is part of a medical device convenience kit	DDM, DFE, KIT, LST	DI, IRC, LWC, PM#	
081.004	foreign manufactured device that is Part of a medical device convenience kit	KIT, DEV, DFE, LST	PM#, DI, LWC;IRC	
081.005	Device constituent part for drug-device combination product	DEV, DFE, LST	DA, IND	
140.000	Import of a device for charity	DEV, DFE, LST	DI, IRC, LWC, PM#	
155.010	Component for further manufacturing into a finished medical device	CPT		LST, PM#
155.011	Device component for use in a drug-device combination product	CPT	DA, IND	
170.000	Repair of medical device and re-exportation	DDM, IFE	DFE, DI, LST, IRC, LWC, PM#	
180.010	Import of research or investigational use in vitro diagnostic device			
180.100*	<ul style="list-style-type: none"> Import of a device for non-clinical use/bench testing Import of device sample for customer evaluation 			
180.200*	Import of a medical device for clinical investigational use	IDE		
920.001	Import of a device that is US goods returned for refund/overstock (to manufacturer)	DDM, LST	DFE, DI, IRC, LWC, PM#	
920.002	Import of device that is US goods returned for sale to a third party	DFE, DDM, LST	DI, IRC, LWC, PM#	
950.001*	Import of a single-use device for domestic reprocessing	DDM, LST	DFE, DI, IRC, LWC, PM#	
950.002*	Import of a multi-use device for domestic reprocessing		DDM, DFE, DI, IRC, LST, LWC, PM#	
970.001	Import for Export: <ul style="list-style-type: none"> Import of a medical device for further processing and re-exportation Importation of a medical device or accessory for further manufacturing into an export-only medical device 	DEV, DFE, IFE, LST		
970.002	Import for Export: <ul style="list-style-type: none"> Importation of a medical device component for further manufacturing into an export-only medical device 	IFE, CPT, DDM, LST		
110.000* 100.000* 940.000* 081.006	<ul style="list-style-type: none"> Public Exhibition/Trade Show Device For Personal Use Compassionate Use/Emergency device Import under enforcement discretion provisions 			

¹: The conditional affirmations are required if applicable to the product being declared. For example, if the product requires premarket clearance (510(k)), then PM# must be provided.

*Annotates that additional information may be needed at time of entry in order for FDA to make a final admissibility decision.

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

General Remarks

If submitting general comments then use the Remarks Type Code = GEN (General Remarks). This is a Free form text relevant to the shipment or the commodity. (68 alphanumeric characters allowed).

Line Value

The value associated with the PGA line number in whole dollars.

All Levels of Packaging (PCS must be base unit)

For Medical Device, this is a mandatory PGA input record that provides FDA with data pertaining to Packaging Qualifier, Quantity and Unit of Measure. This record can be repeated up to six (6) times, once for each unique packaging level. The first record is used to describe the largest (outermost) container and the number of containers at this packaging level. The second record is used to describe the contents of the next smallest container. If needed, qualifiers 3-6 are used in a similar manner (largest to smallest container). The final record (base Unit) must describe the actual amount of the product in the smallest container. **Pieces must be declared as the base unit.**

Valid FDA Units of Measure for Packaging Containers

Code	Description
CS	Case
CT	Carton
BX	Box
PK	Package

Valid FDA Units of Measure for the Base Unit (Last Quantity Transmitted)

Code	Description
PCS	Pieces (Count)

Anticipated Arrival Date, Time and Location

This is a mandatory PGA input record that provides data pertaining to the date, time and location (Anticipated Arrival Location Code is Optional) of the anticipated arrival information for all FDA products.

HTS Codes

Tariff Flag Code	Tariff Flag Code Definition
FD1	FDA data may be required 801 (a)
FD2	FDA data Required 801 (a)
FD3	FDA Prior Notice Data may be required 801 (m)
FD4	FDA Prior Notice Data is required 801 (m)

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Flags do not indicate the FDA program

Disclaimer: Code declaring filing does not require a PGA Message Set.

Valid codes are:

A = product is not regulated by this agency

B = data is not required per agency guidance

Codes A and B are NOT allowed if the HTS is flagged as 'Must Be' provided.

Document Imaging System:

FDA utilizes Import Trade Auxiliary Communication System (ITACS) for documents

ITACS provides the import trade community with three functions:

- the ability to check on the status of an entry,
- the ability to submit entry documentation electronically
- the ability to submit goods availability information for targeted shipments electronically.

ITACS may be accessed at <https://itacs.fda.gov>¹ and the presentation provides an overview and walkthrough of ITACS functionality.

Cargo Release Status Notifications:

The Line with the most severe condition will reflect in the FDA overall status

Entry Level Status Codes are:	
Code	Description
01	DATA UNDER PGA REVIEW
02	HOLD INTACT
04	DATA REJECTED PER PGA REVIEW
06	DO NOT DEVAN
07	MAY PROCEED
08	MOVE TO SECURE HLDNG FCLTY
10	DOCUMENTS REQUIRED
11	INTENSIVE - EXAM/SAMPLE

Entry Line Level Status Codes are:	
Code	Description
01	DATA UNDER PGA REVIEW
02	HOLD INTACT
04	DATA REJECTED PER PGA REVIEW
07	MAY PROCEED

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

PGA Line Level Status Codes are:	
Code	Description
01	DATA UNDER PGA REVIEW
04	DATA REJECTED PER PGA REVIEW
07	MAY PROCEED

If PGA Entry Level, PGA Entry Line Level or PGA Line Level Status Code is:	Then a Valid Status Reason Code can be:	Description
2	11	HOLD INTACT
2	12	EXAM/SAMPLE
2	25	ADDITIONAL VERIFICATION NEEDED
2	15	DATA INACCURATE - CONTACT PGA
4	14	DATA REJECTED PER PGA REVIEW
6	21	EXAM DO NOT DEVAN
7	22	MAY PROCEED
7	23	RELEASED
7	24	RELEASED WITH COMMENTS
10	90	ADDITIONAL INFORMATION NEEDED
11	34	EXAM

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

For questions about FDA Medical Device Message Set

Points of Contact

If you have technical questions about the content of this Supplemental Guide, please email FDA at ACE.Support@fda.hhs.gov.

If you have other questions about this Guide or its data samples, please contact:

Sandy Abbott

Director Division of Compliance Systems
Office of Enforcement and Import Operations
Food and Drug Administration
301-796-3240
Sandra.Abbott@fda.hhs.gov

Max Castillo

Business Process Manager, ACE/ITDS Liaison
Food and Drug Administration
301-796-6675
Max.Castillo@fda.hhs.gov

Jessica Aranda

Communications & Program Analyst
Food and Drug Administration
240-402-9914
Jessica.Aranda@fda.hhs.gov

FDA Supplemental Guide:

<http://www.cbp.gov/document/guidance/fda-supplemental-guide-release-16>

The PGA Message Set chapter/implementation guide and its related Appendix PGA can be found on CBP.gov at:

<http://www.cbp.gov/document/guidance/appendix-pga>

The ACE ABI CATAIR – Custom and Trade Automated Interface Requirements:

<http://www.cbp.gov/document/guidance/pgs-message-set>

Appendix V Government Agency Codes:

<http://www.cbp.gov/document/guidance/appendix-v-government-agency-codes>

Appendix R Intended Use Codes for ACE:

<http://www.cbp.gov/document/guidance/appendix-r-intended-use-codes-ace>

Appendix B Valid Codes:

<http://www.cbp.gov/document/guidance/appendix-b-valid-codes>

FDA – Medical Devices – PGA Filer Data Requirements based on FDA Supplemental Guide

Appendix C:

<http://www.cbp.gov/document/guidance/appendix-c-tariff-abbreviations>

Medical Device Product Classification Database

- This database includes:
 - list of all medical devices with their associated classifications, product codes, FDA premarket review organizations, and other regulatory information.
 - <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPCD/classification.cfm>

Firm's Registration and Listing Status

- CDRH maintains a web site with establishment registration and listing data
 - <http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm>
- This data is updated weekly
- This site does not give device listing numbers since this is proprietary information

Medical Device Resources:

Medical devices, March 24, 2011:

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/ucm248321.htm>

Medical and nonmedical radiation-emitting electronic products, September 6, 2011:

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/ucm271180.htm>

Device Advice: Comprehensive Regulatory Assistance

<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/default.htm>

Establishment Registration & Device Listing

<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm>

This material is provided for general informational purposes only. It is not offered as and does not constitute legal advice. NCBFAA makes no representation regarding the completeness or accuracy of this information. Readers of this material should not act or refrain from acting on the basis of this information without seeking appropriate legal or professional advice on the particular facts and circumstances at issue.